

SUPPLEMENTAL / BID BULLETIN -1

This Bid Bulletin is issued to modify or amend items in the Bid Documents. This shall form an integral part of the bid Documents. Please take note of the changes / adjustments for the project, SUPPLY, DELIVERY AND INSTALLATION OF VARIOUS LABORATORY EQUIPMENT (NEWTON FUND / DA BIOTECH) with an ABC of ₱ 2,639,647.00, as follows:

The following inquiries on specifications were clarified and finalized:

- 1. Item number 1- Incubator
 - a. Temperature of 8°C to 70°C as proposed is acceptable to the end-user
- 2. Item number 4- Hotplate with Magnetic Stirrer
 - a. Temperature range of 50°C to 500°C as proposed is acceptable to the end-user
 - b. Stirring range / speed range 100 1500 rpm as proposed is acceptable to the enduser
- 3. Item number 9- Conventional PCR Machine
 - a. 96 well single blocks as proposed instead of the required dual 48 well is not acceptable to the end-user since they need two separate individual blocks for the project.
 - b. Ramp rate or change of temperature if faster than 5°C is acceptable to the enduser.
 - c. Uniformity better than ±0.4°C is acceptable to the end-user.
 - d. "Other Freebies" are consumables that may be provided which are needed for molecular works.
- 4. Item number 10- Microcentrifuge
 - a. The required capacity is at least 18-place rotor, and the proposed 24 capacity is acceptable as long as it will meet the required capacity for 1.5 and 2.0ml tubes.

Reminders to bidders were enumerated as follows:

- Bid is per Item for 14 items listed.
- Payment of bidding documents is required before submission. Deadline of bid submission is on November 18, 2020 at 1:00 pm, late bids will not be accepted.
- The BAC is still requesting prospective bidders to submit three (3) sets of bidding documents for simultaneous opening and evaluation of the BAC members and TWG for the documentary requirements of the project.
- Bid opening will be face to face, to be held on November 18, 2020 at 4:00pm at CvSU International House 2.
- For those who are interested to attend the bid opening, prospective bidders are advised to send one (1) representative per company. The University is implementing a health protocol to observe, that is to wear mask and face shield and to present a valid health certificate from Barangay or Rural Health Unit.

For guidance and information of all concerned.

Prepared by:	
PRECIOSA G. ERAÑA BAC Secretary	YVES ROY M. TIBAYAN End-user, Newton Project
Certified correct:	
RENT BETONIO TWG Member, Office and Other Sup ORLANGO B. DELOS REYES, MEN	lG
TWG Chair, Office and Other Supplied Approved:	es and Laboratory Equipment
DAVID L. CERO, Ph. D. Chair, BAC for Goods and Consulting	g Services
Received by the Bidder :	