Republic of the Philippines CAVITE STATE UNIVERSITY



www.cvsu.edu.ph

SUPPLEMENTAL / BID BULLETIN -1

This Bid Bulletin is issued to modify or amend items in the Bid Documents. This shall form an integral part of the bid Documents. Please take note of the changes / adjustments for the project, SUPPLY, DELIVERY AND INSTALLATION OF LABORATORY EQUIPMENT (MUFFLE FURNACE AND VACUUM DRYING OVEN) with an ABC of ₱ 1,200,000.00, as follows:

- 1. Muffle Furnace: Capacity: at least 20 L and Maximum Temperature of 1,200°C.
- 2. 3-Phase, 220 Volts electrical requirement.
- **3.** Vacuum Drying Oven: Capacity of at least 30L, dimension will not matter, with at least 1 additional removable thermoshelf that houses the sensor in controlling the temperature.
- 4. Inclusion of a Vacuum Pump that will match and must be connected to the vacuum drying oven.
- 5. 1 Lot bid and evaluation for the 2 Equipment.
- 6. The BAC is still requesting prospective bidders to submit three (3) sets of bidding documents for simultaneous opening and evaluation of the BAC members and TWG.
- 7. Attach pictures/brochures/pamphlets in support to documents being subject for evaluation.
- 8. Payment of bidding documents is required before submission of bids. Deadline of bid submission is on March 4, 2021; 8:00am, late bids will not be accepted.
- 9. Bid opening will be face to face, to be held on March 4, 2021 at 10:00am at CvSU International House 2.
- 10.Bid submission through Courier system is also allowed. However, the bid documents must be received by the BAC before the deadline of submission of bids.
- 11. For those who are interested to attend the face to face bid opening, prospective bidders are advised to send one (1) representative only per company. The University is implementing a health protocol to observe; that is, to wear face mask and face shield and to present a valid health certificate from Barangay or Rural Health Unit.

For guidance and information of all concerned.

. .

Date

Prepared by:		Welstown
PRECIOSA G. ERAÑA		AGNES B. ALIMBOYOGUEN
BAC Secretary		Project Leader / End-user, Research Center
Certified correct:	RENE B BETONIO	Dental and Laboratory Equipment ber, Medical, Dental and Laboratory Equipment
Approved:	DAVID L. CERO, Ph. D Chair, BAC for Goods a	
Received by the Bidder	:	