



Republic of the Philippines  
**CAVITE STATE UNIVERSITY**  
(CvSU)  
Don Severino De Las Alas Campus  
Cavite, Philippines  
☎ (046-4150010-012)  
[www.cvsu.edu.ph](http://www.cvsu.edu.ph)

### **SUPPLEMENTAL / BID BULLETIN -1**

This Bid Bulletin is issued to modify or amend items in the Bid Documents. This shall form an integral part of the bid Documents. Please take note of the changes / adjustments for the project, **Supply, Delivery, and Installation of Various Laboratory Equipment for CVMBS** with an **ABC of ₱ 2,580,000.00**, as follows:

A. Revised Specifications:

**Vertical Autoclave with Programmable Temperature Control System (2 units)**

Design temperature range: 135°C to 140°C  
Power consumption: up to 4 kW  
Safety Device: Overheat (Low Water) Cut-Off Switch Safety Valve and Release Valve or Automatic (once the water went low it automatically switched off) with safety valve but no release valve  
Product demonstration can be done thru video conferencing  
Product demonstration shall be done before product acceptance

**Laboratory Freezer, Temp Set Point: -80 Degrees (1 unit)**

Dimension should conform with the manufacturer standards  
Temperature control range: -86°C to -40°C  
Product demonstration can be done thru video conferencing  
Product demonstration shall be done before product acceptance

**Laboratory Incubator (2 units)**

Incubator temperature: room temp. +5°C to at least 60°C  
Capacity: 100L to 115L  
Power voltage: 220V or 230V  
Product demonstration can be done thru video conferencing  
Product demonstration shall be done before product acceptance

**Cell Line Incubator with CO2 Tank (1 unit)**

Incubator type: Air-jacketed  
The machine should fit table space of not more than the approximate dimension of 16 inches (W) x 19 inches (D) x 21.5 inches (H)  
Controller Display: LED or LCD  
Chamber dimension: at least 12.5 inches (W) x 13.7 inches (D) x 14.5 inches to 15.6 inches (H)  
Product demonstration can be done thru video conferencing  
Product demonstration shall be done before product acceptance

**Liquid Nitrogen Tank, 35l, Low Pressurize, with Casters/Gun (2 units)**

Brand new not reconditioned


**Trinocular Microscope with Camera Attachment (USB) (10 units)**


Product demonstration can be done thru video conferencing  
Product demonstration shall be done before product acceptance

B. Other Concerns and Reminders:


1. The BAC is still requesting prospective bidders to submit three (3) sets of bidding documents for simultaneous opening and evaluation of the BAC members and TWG.
2. Payment of bidding documents is required before submission of bids. Deadline of bid submission is on **March 24, 2022; 8:00am**, late bids will not be accepted.
3. Bid opening will be face-to-face, to be held on **March 24, 2022, at 10:00am** at CvSU International House 2.
4. Bid submission through Courier system is also allowed. However, the bid documents must be received by the BAC before the deadline of submission of bids.
5. For those who are interested to attend the face-to-face bid opening, prospective bidders are advised to send one (1) representative only per company. The University is implementing a health protocol to be strictly observed.


Prepared by:

  
**AL EUGENE L. TORRES**  
Member, BAC Secretariat


  
**EMMANUEL D. MAGO**  
End-User, CVMBS

Certified correct:

  
**LANI S. RODIS**  
TWG Member, Medical, Dental, and Laboratory Equipment

  
**RENE B. BETONIO**  
TWG Chair, Medical, Dental, and Laboratory Equipment

Approved:

  
**DAVID L. CERO, PhD**  
Chair, BAC for Goods and Consulting Services

Received by the Bidder : \_\_\_\_\_  
Date : \_\_\_\_\_