

## Republic of the Philippines CAVITE STATE UNIVERSITY Don Severino de las Alas Campus

Indang, Cavite, Philippines (046) 889-6373 www.cvsu.edu.ph

## INVITATION TO SUBMIT QUOTATION Supply and Delivery of Laboratory Supplies and Equipment for Likas Project

 The Cavite State University (CvSU) invites interested firms/supplier to submit quotation for the project "Supply and Delivery of Laboratory Supplies and Equipment for Likas Project" with an Approved Budget for the Contract (ABC) Nine Hundred Eighteen Thousand Seven Hundred Ninety-Nine Pesos and 90/100 Only (PhP 918,799.90). Quotation received in excess of the ABC shall be automatically rejected at the opening.

Item No.	Quantity	Unit	Description	Unit Cost	Total Cost
1.	1	Unit	MICROSCOPE CAMERA FOR STEREO MICROSCOPE, and also compatible to compound microscope Specifications - Image sensor: 1/2.8 inch - Recordable pixel: up to 1920 x 1080p - Lens mount: C-mount - Live display mode: up to 1920 x 1080p at 30 fps - Exposure: Both Manual and Automatic Control with automatic image correction ISO sensitivity: Standard (equivalent to ISO50) - Interface: with different ports for USB, HDMI, SD Card and stand-alone camera - with built in system that allows measurement of specimens, distance calculation and scale bar display Inclusions: AC Adapter, power chord - Accessories: Monitor	140,000.00	140,000.00
2.	1	Unit	MICROSCOPE, compound Specifications Body: - Trinocular tube model - Stage height: 130 - 135mm - with field diaphragm - with blue light cut filter to prevent eye strains - Digital status display Body tube: - Pupillary distance: 50-70mm - Eye point height: Adjustable - Built-in 0.55x relay lens - Eyepoint height: adjustable to at least 2 positions Ocular: - Eyepiece: 10X with diopter adjustment	250,000.00	250,000.00

- Nosepiece: Reverse type that can handle atleast 5 objectives - Condenser: Vertically movable and centerable - Objectives: 4 objectives (4x, 10x, 40x, 100x oil) Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.  Illumination:	
- Condenser: Vertically movable and centerable - Objectives: 4 objectives (4x, 10x, 40x, 100x oil) Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
centerable - Objectives: 4 objectives (4x, 10x, 40x, 100x oil) Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
- Objectives: 4 objectives (4x, 10x, 40x, 100x oil) Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
oil) Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
Observation Method: - Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
- Brightfield observation method (upgradable to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
to darkfield, phase contrast, simple and color polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
polarizing, and diascopic florescence) Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
Focus: - Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
- Coarse rotation: at least 37.7mm. Min. reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
reading: 2um - Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
- Fine rotation: at least 0.2mm Min. reading: 2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
2um - with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
- with mechanisms to prevent the risk of over focusing, slide breakage and damaging the objectives.	
focusing, slide breakage and damaging the objectives.	
objectives.	
- with automatic light management	
mechanism for uniform brightness and	
intensity regulation.	
- Lamp shelf-life: 50,000 – 60,000 hours	
Stage:	
- Rectangular stage w/ vernier calibrations	
- up to 2 specimen slide holder	
Other Feature: with anti-mold technology	
Inclusions: Microscope cover and power	
cord	
MICROSCOPE, stereo	
Specifications	
- Optical system: Zooming type	
- Magnification: 6.7x - 50x	
- Zoom ratio: 7.5:1	
- Zooming range: 0.67x – 5x	250,000.00
- Tube: built-in C-mount 0.55x magnification	
3.   1   Unit   lens   250,000.00   250,0	
- Eyepiece: 10x; diopter adjuster with clamp	
- Stand: with built-in illuminator and	
capability for episcopic and diascopic	
observation	
- Other Feature: with anti-mold and anti-	
electrostatic technology	
- Inclusions: Microscope cover	
LABORATORY OVEN	
Specifications:	
- Temperature Range: Ambient Temp +70°C	
/ 250°C	
- Temperature Set and Display Sensitivity:	
1°C	150,000.00
4.   2   Units   10   150,00	
- Temperature Fluctuation: ±1°C	
- Timer: 1 minute – 99,9 hours + Hold	
Position	
- Delayed Start Timer: 1 minute – 99,9 hours	
- Number of Shelves (Standard/ Max): 2/7	

			REFRIGERATOR		
			Specifications  10 out # Side by Side HD Invertor Multi		
			- 19 cu. ft Side by Side HD Inverter Multi Door Ref		
			- Energy Efficient Inverter Compressor		
			- Digital Temperature Control with LCD		
_	_	1.1	Display	50,000,00	F0 000 00
5.	1	Unit	- 2 Bottom Freezers	50,000.00	50,000.00
			- Independent Temperature Control		
			- Quick Freeze and Quick Chill Mode		
			- No-Frost - Child Lock		
			- Eco-LED Lights		
			- Spill-Proof Tempered Glass Shelves		
6.	2	bottle	- Adjustable Leveling Feet  ACETONE, 2.5L / bottle	3,000.00	6,000.00
7.	300	bootle	BRISTOL BOARD, 11.5 inches x 16.5 inches	50.00	15,000.00
8.	2	bottle	DENATURED ALCOHOL, 1 Liter / bottle	1,250.00	2,500.00
9.	2	bottle	ETHANOL (ABSOLUTE), 2.5 L / bottle	3,000.00	6,000.00
10.	2	pcs	FORCEPS, fine tip, 6 inches	250.00	500.00
11.	20	pcs	FUNNEL TRAPS	120.00	2,400.00
			GLASS JARS, 470ml cap. with lid		
			Specifications:		
12.	26	Doo	- Material: Glass	202.65	7 274 00
12.	26	Pcs	- Lid type: Metal plastisol lined	283.65	7,374.90
			- Lid color: White / Black		
			- Shape: Straight sided, round, wide mouth		
13.	5	box	GLOVES, 100 pcs / box	125.00	625.00
14.	10	pcs	INSECT BOX, WOODEN 450mm x 300mm	1,200.00	12,000.00
15.	10	pack	INSECT PINS #2, 100 pcs / pack, stainless	750.00	7,500.00
16.	20	pcs	LEG BANDS	110.00	2,200.00
17.	2	pcs	MIST NETS	1,100.00	2,200.00
18.	1	unit	RULER, 40 inches, stainless steel	500.00	500.00
			SPECIMEN BOTTLES, 10 glass vials / set,		
	40	Sets	with cap	250.00	10,000.0080
			Specifications:		
19.			- Material: Glass		
			- Cap material: PP (Polypropylene)		
			- Cap liner material: PE (Polyethylene)		
			- Color: Clear		
			- Capacity: 10ml (0.67 oz)		
	80	Pcs	VIALS, with cap		
			Specifications:		4 000 00
20.			- Material: Glass	50.00	
			- Cap material: PP (Polypropylene)	50.00	4,000.00
			- Cap liner material: PE (Polyethylene)		
			- Color: Clear		
			- Capacity 20ml (0.67 oz)	AL AMOUNT	918,799.80
	J 10,1 JJ.00				

- 2. Delivery Period: \_\_\_\_ calendar days from the receipt of P.O.
- 3. Price quotations must be valid for a period of sixty (60) calendar days from date of submission and shall include all taxes, duties and/or levies payable. Bidders shall also indicate the brand and model of the items being offered.

- 4. Warranty shall be for a period of six (6) months for supplies and materials. Warranty for equipment must not be less than one (1) year from the date of acceptance and shall be accompanied with Warranty Certificate.
- 5. The quotation must be submitted to the Procurement Office through mail, fax or email at the contact details listed below **on or before 5:00 PM** of **May 6, 2024.**

Address : Procurement Office, Administration Building

Cavite State University

Indang, Cavite

E-mail : procurementoffice@cvsu.edu.ph / rfqmain@cvsu.edu.ph

Telefax : (046) 889-6373

6. The CvSU reserves the right to reject any or all quotations and/or proposals and waive any formalities/informalities therein and to accept such bids it may consider as most advantageous to the agency and to the government. CvSU neither assumes any obligation for whatsoever losses that may be incurred in the preparation of bids, nor does it guarantee that an award will be made.

ROSELYN M. MARANAN

BAC Secretary, Goods and Consulting Services