

BID AND AWARDS COMMITTEE Goods and Consulting Services

SUPPLEMENTAL / BID BULLETIN -1

This Bid Bulletin is issued to modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents. Please take note of the changes/adjustments for the project, **Supply**, **Delivery and Installation of Various Laboratory Equipment for the University** (*Remaining from EPA*) an **ABC** of ₱2,100,000.00, as follows:

A. Revised Specifications:

- For Item 1 Refrigerated Centrifuge (CVMBS)
 - Temperature Setting: -20 to 40°C
 - o Omit Cools up to 4°C in less than 10 minutes
 - Size and Weight as per manufacturer's specifications
- For Item 2 UV Cleaner Box for PCR (CVMBS)
 - Stick to Tabletop
 - Walls material Rear: stainless steel or powder coated steel
 - Sides: steel with chemical resistant powder coating or polycarbonate, clear or glass
 - Working surface material will stick to stainless steel
 - o Working Area: Maximum 700 x 550 mm
 - o Dimension will fit to the said area
 - Power Supply will remain AC100-240V 50/60Hz
- For Item 3 Digital pH Meter with pH Pen Tester for liquid solutions (CVMBS)
 - o pH Range 0 to 14.00
 - Resolution 0.001pH
 - Accuracy+0.02
 - Calibration Points Up to 3 (USA & NIST)/Up to 3 (DIN)
 - Temperature Range 0.0 to 100°C
- For Item 4 Nephelometer (CVMBS)
 - No changes made
- For Item 5 Compensated Calorimeter (CAFENR)
 - o No changes made

- For Item 6 Refrigerated Centrifuge (CAFENR)
 - o Corrected to Digital Centrifuge

Speed 200~4000rpm

Timer 0~30 min. digital timer Motor DC Brushless motor

Display LED digital display for time, speed and RCF

R.C.F. 2,325g Max.

Capacity 6x50ml

Standard accessory 50ml tube adapter AC 100~240V 50/60Hz Power supply

The TWG and End-User reiterate that the specified specifications are just the minimum requirements. Prospective bidders can offer higher specification but should be within the specified ABC of the project.

B. Other Concerns and Reminders:

- 1. The award of the project is on per item basis.
- 2. The standard delivery period is set to sixty (60) calendas days.
- 3. The BAC is requesting prospective bidders to submit three (3) sets of bidding documents (Original Copy, Copy 1 and Copy 2) for simultaneous opening and evaluation of the BAC members and TWG.
- 4. Bid documents should contain "ear tags" for easy scanning of all the BAC members.
- 5. Payment of bidding documents is required before submission of bids. The deadline for bid submission is on August 13, 2025; 12:00Noon; Procurement Office, Administration Building, Cavite State University, Indang, Cavite, late bids will not be accepted.
- 6. Bid opening will be face-to-face, to be held on August 13, 2025; at 3:00 PM; CvSU Board Room, Cavite State University, Indang, Cavite.
- 7. Bid submission through courier system is also allowed. However, the bid documents must be received by the BAC before the deadline for submission of bids.
- 8. For the payment of bid documents, the prospective bidders are requested to coordinate with the BAC Secretariat. Online payment through Landbank Link. Biz is accepted.
- 9. For those who are interested in attending the face-to-face bid opening, prospective bidders are advised to send one (1) representative only per company.

Prepared by:

Member, BAC Secretariat

Chair, BAC Secretariat

Noted by:

₽MMANUEL R. MAGO

End-User

End-User

Certified correct:	
RENEB. BETONIO	LANI S. RODIS
TWG Chair	TWG Member
Medical, Dental and Laboratory Equipment	Medical, Dental and Laboratory Equipment
MA. CORAZON V. HERRERA TWG Member Medical, Dental and Laboratory Equipment	ARLEEN C. PANALIGAN TWG Member Medical, Dental and Laboratory Equipment
Smother	EMMANUEL R. MAGO
ANNIÉ MOJICA TWG Member	TWG Member
Medical, Dental and Laboratory Equipment	Medical, Dental and Laboratory Equipment
Medical, Dental and Laboratory Equipment	Medical, Dental and Laboratory Equipment
Approved: BETTINA JOYCE P. ILAGAN, PhD Vice Chair, BAC for Goods and Consulting Services	
Received by the Bidder :	
Date :	